

FEB 10 2010  
2011 (an)

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster cdm Breath Test Instrument*  
*State of Alaska*  
*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130136</u> ✓
Name <u>Daniel G. Gunderson</u>	ID# <u>6602</u> ✓	Date <u>2-6-11</u>
A Agency <u>AST</u>		Phone # <u>543-2294</u>
Instrument Location <u>BETHEL AST</u>		
B Alco S/N <u>X124623</u> ✓ Target Value <u>.081</u> ✓ High Pressure <u>1110</u>		
Alco Test Values <u>.083</u> ✓ <u>.083</u> ✓		
1 <sup>st</sup> Alco 2 <sup>nd</sup> Alco		
Signature <u>Daniel G. Gunderson</u> <u>2-6-11</u>		
(OVER)		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 17<sup>th</sup> day of MAR, 2011 an

Daniel L. Nicholas  
(Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office

BT9 06/03



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

~~X124623~~ <sup>dyn</sup>  
130136 2/10/11

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130136

FEBRUARY 06, 2011

OPERATOR'S NAME:

GUNDERSON/DANIEL/G

OPERATOR'S NUMBER: 6602

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

TEST

O.L. #: 00

DEPT/AGENCY: BET0

CASE/REPORT: 00

TEST TYPE: U

ALCO TARGET VALUE: .061

ALCO S/N: X124623

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.93 in

ALCO TARGET .001 16:54

BLANK TEST .000 16:56

INTERNAL STANDARD VERIFIED 16:56

ALCO TV 29.94 in .003 16:56

BLANK TEST .000 16:57

SUBJECT SAMPLE .000 16:57

BLANK TEST .000 16:58

ALCO TV 29.94 in .003 16:58

BLANK TEST .000 16:59

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130136

FEBRUARY 06, 2011

TIME 17:02

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 43c

BAROMETER: 29.94 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF6  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefgijklmnop  
qrstuvwxyz{|}~!@